

丁腈检查手套

NITRILE GLOVES

不易破 | 强韧性 | 耐腐蚀 | 防滑微麻面

证件齐全 · 品质保证





根据手型选码


测量方法：手掌伸直，大拇指与食指连接处测量至手掌边缘，得出掌宽，对应选码。

手宽<8cm 请选择S号

手宽8-9cm 请选择M号

手宽9-10cm 请选择L号

掌宽



使用步骤 简单易用

4步轻松穿戴，方便省事

USE STEPS ARE SIMPLE AND EASY TO USE



1、打开包装顶部切口



2、抽出一一次性手套



3、穿戴手套(不分正反)



4、拉平手套，完成

HSK 佛山市华士康医疗用品有限公司位于广东佛山市南海区松岗工业区成立于2005年，
是专业研发、生产医疗用品的企业。



全国热线：400-003-9189

单位名称：佛山市华士康医疗用品有限公司

地址：佛山市南海区狮山镇松岗石塘路

电话：0757-85229189

Hsk Medical Apparatus Foshan China Co.,Ltd

Nitrile Examination Gloves - Product Inspection Report

Inspection Report NO.HSK-SYT-Z2020

Name	Nitrile Examination Gloves	Typ & Spec	Fingertip Smoot, Powder Free, L, 100pcs/box		
Origin of Sample	Packaging Workshop	Qty	100 pcs	Lot Number	20200302
Inspection Reference	Yue Chan Medical Device Record NO. 20140029			Date of Production	20200302
Sample Date	20200302	Test Date	20200302	Date of Report	20200303
NO.	Test	Standard		Test Result	conclusion
1	Appearance	The surface should be flat and clean without defects such as stains, macula, damage.		P	P
2	Dimension	Minimum length (mm) : XS:220, S:220, M:230, L:230		250	P
		Width(mm±5): XS:80, S:80, M:95, L:110		115	P
		Minimum thickness(mm): Fingertip Smooth: 0.08 Fingertip Textured:0.11 Maximum thickness(mm): Fingertip Smooth:2.00 Fingertip Textured:2.03		0.11 2.05	P
3	Before Aging	Breaking strength, min: 7.0N Elongation at break, min: 500%		8.0N 520%	P
4	After Aging	Breaking strength, min: 7.0N Elongation at break, min: 400%		7.3N 410%	P
5	Impermeability	According to the appendix A, there should be no leakage beyond 40mm from the sleeve.		P	P
Inspection Conclusion		The inspection results for the inspected items are in accordance with the relative specifications in the inspection reference, and the corresponding performance of the inspected samples is qualified.			

Composed by:Zao jie

Examined & Approved by:Yuncheng Wang



CERTIFICATE 合格证

DESCRIPTION OF GOODS产品名称	NITRILE EXAMINATION GLOVES(NON-STERILE) 丁腈检查手套（非灭菌）
BRAND 品牌	HSK
SIZE 尺寸	M
MAIN MATERIAL	NITRILE 丁腈
STANDARD 执行标准	GB 10213-2006
QUANTITY 数量	100 PCS
MFG.DATE 生产日期	2020.04.03
LOT NO. 生产批号	20200403
EXP. DATE 保质期	20220402
MANUFACTURER 生产商	HSK MEDICAL APPARATUS FOSHAN CHINA CO., LTD. 佛山市华士康医疗用品有限公司
ADDRESS 生产地址	Shitang Road, Shishan Town, Nanhai District, Foshan, Guangdong, China 佛山市南海区狮山镇松岗石塘路
INSPECTOR 检验员	C-03